



SUBJECT: Advoque Safeguard / Customers and Distributors

Friday, August 5, 2022

RE: Status – Advoque Safeguard N95 Particulate Filtering Respirator – NIOSH TC-84A-9284

We are distributing this notice with regard to the Advoque Safeguard N95 Particulate Filtering Respirator, model number ASG100.

Advoque Safeguard is voluntarily withdrawing its NIOSH certification for model number ASG100, NIOSH TC-84A-9284 as of August 5, 2022.

Thus, the ASG100 can no longer be manufactured, assembled, labeled, sold, or distributed as a NIOSH approved product.

Any ASG100 units that were used prior to August 5, 2022 would have been NIOSH approved at the time of usage.

Advoque Safeguard is taking this action given a third party test performed on one of its lots that recorded some of the units with inconsistent filtration performance to the 95% filtration standard. While all of the lots for ASG100, inclusive of this lot, were successfully tested prior via a prominent third party testing lab as part of the Advoque Safeguard production protocol, the incongruous nature of the recent results have predicated Advoque Safeguard to take this action.

Please reach out to us with any questions you may have at qc@advoquesafeguard.com